

Guidelines on Monitoring of Airborne Chemical Hazardous to Health 2022

ERRATA

The following list provides a description of the changes made to the publication since the original version was printed.

CHAPTER 3: SAMPLING TECHNIQUE

1. Page 17, (i) Sorbent Tubes, after fifth sentence:
add new sentence "Some method for example US NIOSH uses the following formula to identify breakthrough:

Breakthrough and possible sample loss is occurred if $W_b > \frac{W_f}{10}$ where W_f and W_b is the weight of the front and back portion of sorbent sections respectively."

CHAPTER 4: MONITORING STRATEGY

2. Page 28, Example: Number of Sampling Media Calculation, first sentence a):
Replace "0500" *with* "0501"

CHAPTER 6: SAMPLING RESULTS MANAGEMENT AND INTERPRETATION

3. Page 35, c) Breakthrough Identification, after third sentence:
Add new sentence "For breakthrough identification, HT should be able to clearly identify method for breakthrough identification."

Date : 18 July 2023